



# Test Report

No.: T32020270589SN

Date: SEP 22, 2020

Page 1 of 2

TREND WEALTH INDUSTRIAL LIMITED  
UNIT B, 1/F., EASTERN INDUSTRIAL BUILDING, 42-50 KWAI TING ROAD, KWAI CHUNG,  
NEW TERRITORIES, HONG KONG

The following samples were submitted and identified on behalf of the client as:

### 3-PLY DISPOSABLE FACE MASKS

SGS Case No. : CA320202731749  
 Lot No. / Batch Code : NOT PROVIDED  
 Sample Description : BLUE MASK  
 Quantity Submitted : 20 PCS.  
 Manufacturer : TREND WEALTH INDUSTRIAL LIMITED  
 Country of Origin : HONG KONG  
 Sample Receiving Date : JUL 23, 2020  
 Test Performing Date : JUL 23 TO SEP 22, 2020

Test Requested : Please refer to the result summary.

Test Method & Results : Please refer to next page(s).

Result Summary :

Test Requested	Conclusion
Viral filtration efficiency (VFE) (With reference to ASTM F2101)	See Result

\*\*\*\*\*

Signed for and on behalf of  
SGS Hong Kong Ltd.

Au Kam Chi, Gigi  
Technical Manager

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



# Test Report

No.: T32020270589SN

Date: SEP 22, 2020

Page 2 of 2


## Viral filtration efficiency (VFE) With reference to ASTM F2101-19

Test Side : White Side (Inside)  
 Pre-Conditioning : Minimum of 4 hours at 21±5°C and 85±5% R.H.  
 Test Condition : 21±2 °C and 30-50% R.H.  
 VFE Test Area : ~ 40 cm<sup>2</sup>  
 VFE Flow Rate : 28.3 Litre per minute (L/min)  
 Mean Particle Size : 2.8 µm  
 Positive Control Average : 1.2 x 10<sup>3</sup> PFU  
 Negative Monitor Count : <1 PFU

Test Specimen	Percent VFE (%)
1	99.6
2	99.7
3	99.7
4	99.0
5	99.4

Note : The above test results were conducted by a SGS assessed competent subcontractor laboratory

### Photo Appendix



\*\*\* End of Report \*\*\*

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.